



Type examination certificate

No. CH-MI004-09011-00

Bauartprüfzertifikat

Applicant:
Auftraggeber:

G2 MISURATORI SRL
Via san Martino 38
14100 Asti
Italy

Requirements:
Anforderungen:

Swiss ordinance on measuring instruments (SR 941.210) of February 15, 2006 annex 2 module B;

Directive 2004/22/EC of the European Parliament and Council of March 31, 2004 on measuring instruments (MID), appendix B, class of measuring instruments MI-004.

Conformity of superseded standard:
EN 1434: Edition 2007

Type of instrument:
Geräteart:

Single-jet impeller flow sensor as a subassembly of a heat meter

Type designation:
Typenbezeichnung:

CSCUL

Accuracy class(es):
Genauigkeitsklasse(n):

2

Characteristics:
Kenndaten:

PN 16, DN 15 / 20, $T_{\max} = 90$ °C, Class A, E1, $q_p 1.5 / 2.5$

Certificate valid until:
Zertifikat gültig bis:

October 29, 2019

CH-3003 Bern-Wabern, October 30, 2009

Notified body:
Benannte Stelle:

Certificate authority METAS-Cert
No. 1259

For the test:
Für die Prüfung:

Dr. Hugo Bissig

Jürg Ramseyer, Head of METAS-Cert